

Conformity of test results

GSI-PL FB001


This statement of conformity applies in addition to the GTC (General Terms and Conditions) and respectively to the customer order.

According to the standard DIN EN ISO/IEC 17025:2018, chapter 7.8.3.1 and 7.8.6, there is a requirement to agree on the conformity of test results with the customer. The applied decision rule must be documented.

GSI mbH uses the following decision rules:

- 1. If the decision rule is specified in the standards or specifications of the testing commissioned, these shall be deemed to have been agreed with the customer.
- 2. If the customer requires a different decision rule or his own requirement for the test result, he must notify them separately in writing with the order request / order and indicate the corresponding decision case (No. I VII) according to this document.
- **3.** Unless the above items 1 and 2 are not present, the following decision rule is applied in principle.

Decision rule according to 3. (as mentioned above)



The conformity statement will not take into account measurement uncertainties. The requirement is fulfilled if the measured value is less than or equal to the tolerance limit (decision case IV).